

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JAN 1 2011
2011
3/9/11

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130122</u> ✓	
Name	<u>Kirsten Hansen</u>	ID#	<u>5440</u> ✓
		Date	<u>01/11/2011</u> ✓
A	Agency	<u>Alaska State Troopers / DJT0</u>	Phone # <u>(907) 895-4800</u>
	Instrument Location	<u>Delta Jct. Court House</u>	
B	Alco S/N	<u>75602</u> ✓	Target Value <u>.081</u> ✓
		High Pressure	<u>850</u>
	Alco Test Values	<u>.082</u> ✓ 1 st Alco	<u>.081</u> ✓ 2 nd Alco
	Signature	<u>Kirsten Hansen</u>	
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 11th day of MAR, 2011 DNW

Daniel L. Nicholas (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

BT9 06/03



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130122

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130122

JANUARY 11, 2011

OPERATOR'S NAME:

HANSEN/KIRSTEN/Z

OPERATOR'S NUMBER: 5440

SUBJECT'S LAST NAME:

JAN

SUBJECT'S FIRST NAME/MI :

VOC/11

O.L. #: 2011111

DEPT/AGENCY: DJT0

CASE/REPORT: 110111

TEST TYPE: U

ALCO TARGET VALUE: .001

ALCO S/N: 75602

— BREATH ANALYSIS —

.001 ADJUSTED FOR 29.88 in		
ALCO TARGET	.000	13:59
BLANK TEST	.000	14:00
INTERNAL STANDARD	VERIFIED	14:00
ALCO TU 29.88 in	.002	14:00
BLANK TEST	.000	14:01
SUBJECT SAMPLE	.000	14:02
BLANK TEST	.000	14:03
ALCO TU 29.88 in	.001	14:03
BLANK TEST	.000	14:04

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130122

JANUARY 11, 2011

TIME 14:16

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	40c
BREATH TUBE:	42c
BAROMETER:	29.88 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~